Knee Immobilizer

Application, removal, and care guidance for your immobilizer



APPLICATION

- Open the knee immobilizer and remove the stay assemblies.
- Position the knee immobilizer with the wide end towards the thigh and the back stays along the midline of the leg.
- Use the small hook tab to hold the body around the knee.
- Attach stay assemblies, inside first, to accommodate calf and thigh circumference.
- Starting at the bottom and moving up, pull straps forward and around the leg, attaching straps where comfortable.

REMOVAL

- Open hook-and-loop closures and unwrap the splint.
- Pull gently away from the leg.
- Re-attach the hook-and-loop closures to increase the life of your product.

CARE

- Remove metal stays from both sides of the immobilizer.
- Close all contact closures.
- Hand wash in warm water with mild soap.
- Product should be air-dried; do not use high temperatures to wash or dry.

> **Note:** *If you experience persistent pain, swelling, numbness, or discoloration while wearing this device, remove it and contact your healthcare provider.*